



CERTIFICATE OF ANALYSIS

cfDNA Reference Standard Set

IMPORTANT: Centrifuge tube prior to opening to prevent loss of product

PRODUCT INFORMATION			
Set Catalogue No.	HD780	Serial No.	16922
Date of Manufacture	Aug 2015		
Description	Multiplex I Cell Free DNA Reference Standard Set		

PROPERTIES	
Quantity	4 vials of 350ng
Concentration	40ng/ μ L
Fragmentation	Fragmented DNA (average size 160bp)
Buffer	Tris-EDTA (10 mM Tris-HCl, 1 mM EDTA), pH 8.0
Storage Conditions	+4°C
Use By	Stable for 12 months from date of manufacture (as supplied)

QUALITY CONTROL		
TEST	TEST METHOD	PASS/FAIL
Allelic Frequency	Droplet Digital™ PCR <ul style="list-style-type: none"> 0% expected frequency for wild type, up to 0.01% 0% expected frequency for wild type, up to 0.1% [For targets unverified at 0.1%] For AF \geq 10% \pm 20% For AF \geq 5% < 10% \pm 30% For AF \geq 1% < 5% \pm 40% For AF <1% \pm 50% 	Pass
Fragmentation Size	D1000 DNA ScreenTape assay (Agilent Tech) <ul style="list-style-type: none"> Fragmented DNA of average size 160bp \pm 10% (144bp – 176bp) 	Pass
Quantification	Qubit® dsDNA BR assay (Invitrogen) <ul style="list-style-type: none"> 40.0 ng/μL \pm 10% (36.0 – 44.0 ng/μL) 	Pass

CUSTOMER SUPPORT

Technical	technical@horizondiscovery.com
Orders	orders@horizondiscovery.com

FOR RESEARCH USE ONLY